

Certificado No.: 67368 **Equipo (Instrument):** Electrocardiógrafo
Fecha de Calibración: 2026-03-24 **Marca (Brand):** SCHILLER

Tipo (Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	Incertidumbre (Uncertainty)
ECG	60 BPM	59 BPM	-1 BPM	± 0.89 BPM
ECG	120 BPM	121 BPM	1 BPM	± 0.89 BPM
ECG	180 BPM	179 BPM	-1 BPM	± 0.89 BPM
ECG	240 BPM	239 BPM	-1 BPM	± 0.89 BPM
ECG	300 BPM	298 BPM	-2 BPM	± 0.89 BPM
Sensibilidad	1 mV	1 mV	0 mV	± 0.011 mV
Sensibilidad	1.5 mV	1.5 mV	0 mV	± 0.011 mV
Sensibilidad	2 mV	2 mV	0 mV	± 0.011 mV
Sensibilidad	2.5 mV	2.5 mV	0 mV	± 0.011 mV
Sensibilidad	3 mV	3 mV	0 mV	± 0.011 mV
Sensibilidad	4 mV	4 mV	0 mV	± 0.011 mV
Sensibilidad	5 mV	6 mV	1 mV	± 0.011 mV
Sensibilidad	10 mV	11 mV	1 mV	± 0.011 mV
Sensibilidad	25 mV	26 mV	1 mV	± 0.011 mV