

Certificado No.:

67786

Controlador de Temperatura

Fecha de Calibración:

2026-03-31

FULL GAUGE

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
11:00:00 AM	2.5	3.2	0.7
11:01:00 AM	2.7	3.1	0.4
11:02:00 AM	3.4	3.5	0.1
11:03:00 AM	3.8	3.9	0.1
11:04:00 AM	4.3	4.4	0.1
11:05:00 AM	4.7	4.8	0.1
11:06:00 AM	4.9	5.0	0.1
11:07:00 AM	3.8	4.4	0.6
11:08:00 AM	3.0	3.7	0.7
11:09:00 AM	2.3	3.0	0.7
11:10:00 AM	2.9	3.1	0.2
11:11:00 AM	3.6	3.3	-0.3
11:12:00 AM	4.1	4.1	0.0
11:13:00 AM	4.5	4.5	0.0
11:14:00 AM	4.8	4.9	0.1
11:15:00 AM	4.7	5.0	0.3
11:16:00 AM	3.6	4.3	0.7
11:17:00 AM	2.7	3.4	0.7
11:18:00 AM	2.5	3.0	0.5
11:19:00 AM	3.2	3.2	0.0
11:20:00 AM	3.9	3.8	-0.1
11:21:00 AM	4.3	4.4	0.1
11:22:00 AM	4.6	4.7	0.1
11:23:00 AM	4.9	4.9	0.0
11:24:00 AM	3.9	4.4	0.5
11:25:00 AM	2.9	3.3	0.4
11:26:00 AM	2.2	2.9	0.7
11:27:00 AM	3.0	3.4	0.4
11:28:00 AM	3.7	4.0	0.3
11:29:00 AM	4.2	4.4	0.2
Promedio	3.7	3.9	
Desviación	0.839	0.694	
Max	4.9	5.0	
Min	2.2	2.9	

Certificado No.:

67786

Controlador de Temperatura

Fecha de Calibración:

2026-03-31

FULL GAUGE

